

# Replacement Parts

## Syringes and Power Supplies

### Standard Syringes

Part Number	Syringe Size	Optimal Range
59000-05	10 µL	1 – 10 µL
59000-10	25 µL	2.5 – 25 µL
59000-15	50 µL	5 – 50 µL
59000-20	100 µL	10 – 100 µL
59000-25	250 µL	25 – 250 µL
59000-30	500 µL	50 – 500 µL
59000-35	1.0 mL	100 µL – 1.0 mL
59000-40	2.5 mL	250 µL – 2.5 mL
59000-45	5.0 mL	500 µL – 5.0 mL
59000-50	10.0 mL	1 – 10.0 mL
59000-55	25.0 mL	2.5 – 25.0 mL
59000-60	50.0 mL	5 – 50.0 mL

### SaltLine Syringes

Part Number	Syringe Size	Optimal Range
208335	1 mL	100 µL – 1.0 mL
208336	5 mL	500 µL – 5.0 mL
208337	10 mL	1 – 10.0 mL

### Power Supply & Power Cords

Part Number	Description
61092-01	Power Supply Universal (110–220 VAC)

Part Number	Description	Diagram of Plug
355008	Switzerland	
3892-01	Continental Europe, Russia, Schuko	
3892-02	Australia, New Zealand, Argentina, China	
3892-03	UK, Ireland, Malaysia, Middle East	
3892-05	USA, Canada, Mexico, Central America, Brazil, Japan	

### Selecting a Syringe:

Select the smallest syringe with a maximum volume that is greater than the largest volume to be dispensed. Ideally the smallest volume to be dispensed should fall within the optimal ranges listed to the left. The Microlab 600 can dispense volumes below the optimal range but there will be some impact on accuracy and precision. The SaltLine Syringes should be used when working with solutions that have a high salt concentration. Contact a Hamilton sales representative for additional assistance.

Power Cord



Power Supply

